NATIONAL INSTITUTE OF PUBLIC HEALTH
Šrobárova 48 Prague 10 100 42

YUUKI Company, s.r.o.
Křižanská 1
460 10 Liberec

YOUR REF..:	30.1.2019
DATE:	SZÚ 405/2019
OUR REF.:	CTZB 187- 405/19-37 EX 190163
PROCESSED BY:	RNDr. M. Rucki, PhD.
TEL/FAX.:	267 08 23 76
E-MAIL:	marian.rucki@szu.cz
DATE:	6.3.2019

Subject: **EXPERT OPINION on the health safety of menstrual cups. - COLORFUL**

SUBJECT OF THE REQUEST:
Our decision regarding your request from 30.1.2019 on evaluation of the health safety of menstrual cups under Act No. 102/2001 Sb., as amended, on General Product Safety is as follows:

SAMPLES PROVIDED: **1) Menstrual cup RAINBOW – silicone, colorful**

Manufacturer: **YUUKI Company, s.r.o., Křižanská 1, 460 10 Liberec**

DOCUMENTATION PROVIDED:
Specification of the product provided in the request from 30.1.2019.

TESTING PREFORMED:
Chemical analyses (presence of primary aromatic amines under ČSN 621156, total migration in the water solution measured under ČSN 621156, volume of reducing substances in the water solution under ČSN 621156) and the cytotoxicity assay (under ČSN EN ISO 10993-5:2005, Part 5) were carried out in Test Laboratories no 1206, certified by the Czech Accreditation Institute, in the Centre of Toxicology and Health Safety (CTZB).
The analyses were carried out in the extent of the Methodological Guidelines of the National Institute of Public Health (SZÚ) No. 1/2000 on evaluation of products in direct contact with the human body by skin or by mucosa (AHEM No. 3/2000).

The provided samples were used for the above analyses.

EXPERT OPINION:
The subject of the assessment was the analysis of possible undesirable effects of the material in contact with skin.
The presence of primary aromatic amines in the samples are below the detection limits.
The values of total migration in the water solution of the samples are below the limit for a safe product.
The volume of reducing substances in the water solution of the samples is also below the limit for a safe product.
The results of the cytotoxicity assay showed that the products are not toxic for tissue-culture cells.

Results:

Sample (extract)	Viability (% control)
Rainbow - 10%	101,8
25%	97,1
50%	93,7
100%	98,3
NK	100.0
PK - LS 1 µg/ml	90,1
LS 10 µg/ml	35,5
LS 20 µg/ml	7,4

Assessment based on the results:

Samples Rainbow are not toxic for tissue-culture cells under the test conditions.

Date: 6.3.2019

Analysis performed by: RNDr. K. Kejlová, Ph.D.

CONCLUSION:
Based on the assessment of the provided documentation, chemical analysis and cytotoxicity assay, the decision was reached that the products menstrual cups RAINBOW are safe for human health when in contact with skin. The results of the performed chemical analyses and cytotoxicity assays confirm health safety under Act No. 102/2001 Sb., as amended, on General Product Safety.
This statement applies is valid only for the provided samples and the conclusions drawn from their evaluation can only be applied on like products, whose content and properties correspond entirely to thus evaluated samples.

NATIONAL INSTITUTE OF PUBLIC HEALTH
Centre of Toxicology and Health Safety for Cosmetics
Šrobárova 48, 100 42 Praha 10

MUDr. Dagmar Jírova, CSc.
Director of the Centre of Toxicology and Health Safety

ATTACHEMENTS:
Chemical Analyses Protocol no. CTZB 187-405/19-37
Cytotoxicity Assay Protocol no. CTZB 187-405/19-37

CHEMICAL ANALYSES PROTOCOL

National Reference Centre for Cosmetics
National Institute of Public Health Prague, Centre of Toxicology and Health Safety Šrobárova 48, 100 42 Prague 10

Contracting Entity: **YUUKI Company, s.r.o., Křižanská 1, 460 10 Liberec**

Protocol Number: **CTZB 187-405/19-37**
Date of the test: **25.2.-27.2.2019**

Sample description: **1) menstrual cup RAINBOW – colourful**

(orange – 2327, blue – 1521, red – 1102B, green – 1400)

Manufacturer: **YUUKI Company, s.r.o., Křižanská 1, 460 10 Liberec**

CHEMICAL ANALYSES

Measured parameters:

Determination of primary aromatic amine

Solution: 37 °C ± 2 °C, 24 h., 100 cm² in 100 ml, (2g to 100 ml – for foaming material) method according to CSN 62 1156
device: VARIAN CARY 1E

Volume of reducing substances in the water solution

solution: 37 °C ± 2 °C, 24 h., 8 g in 100 ml, method according to CSN 62 1156

Total migration in the water solution

solution: 37 °C ± 2 °C, 24 h., 100 sq cm in 100 ml, method according to CSN 62 1156

Results of the tests:

sample no.: 1) menstrual cap COLORFUL - silicone				
test	**value measured**	**unit**	**unit**	**limit**
chemical analyses				
Primary aromatic amine	<0,03*	$mg.l^{-1}$	±15%	0,05
reducing substances in the water solution	1,8	ml $KMnO_4$ ($c = 3.10^{-3}$ $mol.l^{-1}$).50ml^{-1}	± 18%	30
total migration in the water solution	1,0	$mg.dm^{-2}$	± 12%	10

*values are lower than limits set for quantification

NATIONAL INSTITUTE OF PUBLIC HEALTH
National Reference Centre for Cosmetics

Chemical analyses were carried out in the extent of the Methodological Guidelines of the SZÚ no. 1/2000 on evaluation of products in direct contact with the human body by skin or by mucosa (AHEM č. 3/2000).

Date: 28.2.2019
Test performed by: RNDr. Marián Rucki, PhD.

NATIONAL INSTITUTE OF PUBLIC HEALTH
National Reference Centre for Cosmetics

CYTOTOXICITY TEST REPORT – METHOD IN VITRO

National Institute of Public Health Prague, Centre of Toxicology and Health Safety
Šrobárova 48,100 42 Prague 10

Test laboratory no. 1206, certified by the Czech Accreditation Institute

Contracting Entity: YUUKI Company, s.r.o., Křižanská 1, 460 10 Liberec

Protocol Number: CTZB 187-405/19-37

Date of the test: 18.2.-20.2.2019

List of samples: VZ 1 - Menstrual cap sample Rainbow, silicone, colorful

The test was carried out according to CSN EN ISO 10993-5:2010 – Biological assessment on medical devices - Part 5: Tests for in vitro cytotoxicity.

Cell line: Mouse fibroblast - line Balb/c 3T3
Culture medium: DMEM containing antibiotics (PNC 100 IU/ml, STM 100 µg/ml) with 10 % of inactive bovine serum
Positive control: Sodium laureth sulphate (SLS)
Negative control: Hydro - poly[(2-hydroxyethyl) methacrylate]
Preparation of the extract: 0.1 g of the sample in 1 ml of the extraction agent (DMEM without serum), 24 hours at 37°C. 100% extracts were then diluted by DMEM without serum.

Procedure: After 24-hour pre-culture, the cellular culture was exposed to 10%, 25%, 50% and 100% extracts of the tested samples VZ1 for the period of 24 h. (37°C, 7.5% CO_2).Then the viability of the cellular culture was determined based on incorporation of vital dye (neutral red) by the fluorometric method. The viability of the culture exposed to the tested sample was then compared to the viability of the negative control.

Cytotoxicity level of the extract:

viability equal and above 70 % .....non-cytotoxic
viability equal and above 50% and below 70 %.....mildly cytotoxic
viability equal and above 30% and below 50 %..... …medium cytotoxic
viability below 30 % ..... strongly cytotoxic

NATIONAL INSTITUTE OF PUBLIC HEALTH
National Reference Centre for Cosmetics